


## pH/mV/°C/ORP Analyzer Model : $\mu$ pHCal<sub>100</sub> (Upto 3 / 5 Point Calibration)



### Back Panel View



Supplied with Combination Glass pH Electrode (0-70°C), Standard Buffer Tablets, Temperature Probe, Electrode Stand with Clamp, Dust Cover, DC Adaptor, Operational & Instruction Manual, Manufacturer Test Certificate and Warranty Card.

### Technical Specification :

pH

**pH Range :** -1.000 to 15.000 pH

**Selectable Resolution :** 0.1 / 0.01 / 0.001 pH

**Relative Accuracy :**  $\pm$  0.002 pH

mV

**Range – mV / RmV :**  $\pm$  2000.0

**ORP :** EH autocalibration to standard hydrogen electrode

**Resolution :** 0.1

**Relative Accuracy :**  $\pm$  0.2 mV or 0.05%, whichever is greater

Temperature

**Temperature Range :** 0 to 130.0°C

**Temperature Resolution :** 0.1°C

**Relative Accuracy :**  $\pm$  0.1 °C

**Backlit LCD Display :** 20 x 2 line Alphanumeric

**Auto Buffer Recognition :** 1.68, 4.01, 6.86 / 7.00, 9.18 / 10.01 & 12.45 pH @25°C

**Buffer Standard :** US / NIST & EURO

**Power Supply :** DC Adaptor

**Salient Features :** Auto Buffer Recognition, Simultaneous Display of pH, mV & Temperature, Parallel & Serial Port for Printer & PC attachment, Report Formation as per GLP Compliance, More than 150 analysis data storage with corresponding Calibration data with Batch Number, Real Date & Time, Operator Name, Sample Name, searching of data as per Batch Number. Calibration Reminder Alarm (0-99 Hrs), Results Hold / Un-Hold Facility etc.

**Optional for  $\mu$ pHCal<sub>100</sub> :** RS-232 to USB interface with Cable & Driver Software to Transfer Data From Instrument to PC

### Error Message :

Bad Buffer / Electrode ■ Temperature Probe ■ pH over range

### Optional

\* NIST Traceable Standard for pH and ORP/Redox

\* IQ/OQ/PQ/DQ Documentation

\* Customized Electrodes

\* GMP Model on Request

\* Calibration Certificate with Tracability

\* Third Arm Stand

```

CALIBRATION REPORT
-----
COMPANY NAME           : ANALAB (VADODARA)
REPORT PRINT DATE & TIME : 13-NOV-2009 - 16:55
CALIBRATION DATE & TIME : 13-NOV-2009 - 16.31
CALIBRATION DONE BY   : AAAAAAAAAAAAAAAAAAAAAA
CALIBRATION POINT     : FIVE
    
```

```

Std. Buffer 1.68          Std. Buffer 7.00/6.86          Std. Buffer 10.01/9.18
-----
pH = 01.680             pH = 07.185             pH = 09.589
mV = +0314.7           Offset= -0011.0         mV = -0153.2
Temp. = 025.0           Temp. = 025.0           Temp. = 025.0
% Slope S1 = 100.09     Buffer used:07.00       % Slope S3 = 110.27
    
```

```

Std. Buffer 4.01          Std. Buffer 12.45
-----
pH = 04.291             pH = 12.451
mV = +0160.2           mV = -0322.5
Temp. = 025.0           Temp. = 025.0
% Slope S2 = 096.77     % Slope S4 = 099.93
    
```

Calibration done on MicropHCal100  
Instrument Serial No. : 200910/704

```

ANALYSIS REPORT
-----
COMPANY NAME           : ANALAB (VADODARA)
ANALYSIS DATE & TIME   : 13-NOV-2009 - 16:40
CALIBRATION DATE & TIME : 13-NOV-2009 - 16.31
BATH NUMBER/ID        : AAAAAAAAAA
SAMPLE NAME            : AAAAAAAAAAAAAAAAAAAAAA
USER NAME              : AAAAAAAAAAAAAAAAAAAAAA
    
```

pH = 04.291  
mV = +0160.2  
Temp. = 025.0

Signature

Analysis done on MicropHCal100.  
Instrument Serial No. : 200910/704